



HELAGO-CZ, s.r.o.  
Commercial Register maintained by the Regional Court in Hradec Králové  
Section C, File 17879  
Kladská 1082  
500 03 Hradec Králové 3  
Company ID: 25 96 39 61, VAT: CZ 25963961  
Phone: 495 220 229, 495 220 394  
Fax: 495 220 154  
GSM gate: 602 123 096  
E-mail: info@helago-cz.cz  
Web: http://www.helago-cz.cz

## MX-50 - Moisture Analyzer

Order code: **0105.MX50**



Cena bez DPH  
Price with VAT

3.543,00 Eur  
4.287,03 Eur

### Parameters

Type Balances

Series

Approval (Verif. ČMI)

Capacity (maximum)

Moisture Analyzers

MF, ML, MS, MX

NO

51 g

Min. weighing value	1
Division [%]	0,01
Interface	RS-232C
Calibration	External
Weighing pan [mm]	Ø 85
Quantitative unit	ks

- Fast and uniform heating with a straight halogen lamp and innovative SRA (Secondary Radiation Assist) technology shorten measurement time
- Standard WinCT-Moisture (for MS-70 & MX-50 only) for easy analyzing mode setting through PC and for real-time graph displaying of moisture content rate change while measuring
- Sodium Tartrate Dehydrate with its stable moisture rate comes standard for accuracy checking
- Optional temperature calibrator, calibration result can be output
- Memory function allows 20 (10 for MF-50, 5 for ML-50) suitable measurement conditions according to sample and 100 (50 for MF-50, 30 for ML-50) measurement result data storage and outputs
- Four choices of measurement programs, Standard, Automatic Complete, Timer, and Manual Mode are selectable
- Clear and easy-to-see, large VFD display shows measurement, setting value, change in moisture content, action status, data number and other important information
- Low maintenance cost assured, thanks to user replaceable halogen lamp without unit downtime with protective chamber for easy clean
- Progress window for heating check
- "Quick Mode" in the measurement program for drying the sample faster by heating the sample at 200°C for 3 minutes.
- Self check function
- Conformity to GLP,GMP,GCP and ISO with date /time ,ID calibration data and check record outputs