

HELAGO-CZ, s.r.o.
Commercial Register maintained by the Regional Court in Hradec Králové
Section C, File 17879
Kladská 1082
500 03 Hradec Králové 3

Company ID: 25 96 39 61, VAT: CZ 25963961

Phone: 495 220 229, 495 220 394

Fax: 495 220 154 GSM gate: 602 123 096 E-mail: info@helago-cz.cz Web: http://www.helago-cz.cz

# VBF.129.60-G - One wing door -

Order code: 8101.VBF12960G



Information about product price on demand

Parameters

Types of safety cabinets

Color

Interior equipment

Quantitative unit

Typ 90 - křídlové dveře

Žlutá

Police

ks

- Safe and approved storage of hazardous materials in workrooms
- 90 minutes fire resistance cabinet in accordance with EN 14470-1 (Type 90)
- Cylinder locking (suitable for masterkey systems)
- The extraction air inlet is flush with the ceiling of the cabinet
- Adjusting aids to compensate for uneven floor
- Earthing link (standard)

## Operation

- Smooth, air-damped door closing
- Minimal effort: door opening easy and safe
- Door closing system outside the storage compartment: increased protection of moving and safety relevant components against corrosion

## Wing doors

• 90° opening, interior of the cabinet completely visible

### Door open arrest system (optional)

- Holds door open and eases the storage and removal inside the cabinet
- Door open arrest system with an ingrated, thermal release system (fusible link), automatic closing of the doors

#### Interior equipment

- 3 adjustable shelves with inclination and pull-out protection, adjustable in different height in 32 mm load on one shelf (uniform distributed) approx. 75 kg
- · Easily removable bottom collection sump with three-sided lip seal for safe collection of leakage
- 1 perforated metal plate incerted at the bottom of collecting sump

## Cabinet with width of 600 mm (VBF.129.60)

- One wing door
- Dimensions w  $\times$  d  $\times$  h (mm): exterior approx. 600  $\times$  615  $\times$  1 298, interior approx. 450  $\times$  520  $\times$  1 070
- Weight: approx. 170 kg
- Bottom collection sump volume 22 I