




HELAGO-CZ, s.r.o.  
Commercial Register maintained by the Regional Court in Hradec Králové  
Section C, File 17879  
Kladská 1082  
500 03 Hradec Králové 3  
Company ID: 25 96 39 61, VAT: CZ 25963961  
Phone: 495 220 229, 495 220 394  
Fax: 495 220 154  
GSM gate: 602 123 096  
E-mail: [info@helago-cz.cz](mailto:info@helago-cz.cz)  
Web: <http://www.helago-cz.cz>

AD-1690 - Leak tester  
Order code: **0109.AD1690**





Cena bez DPH

1.617,00 Eur

Price with VAT

1.956,57 Eur

Parameters

Quantitative unit

ks

- Detailed description
- Quick testing of less than 5 seconds
- Leak detection by depressurization ensuring no dust enters the object being measured
- Replaceable air filter to protect the AD-1690 from dust when it intakes air
- RS-232C interface to output data to a PC or an A&D peripheral device
- Three types of attachments to cover micropipettes with capacities ranging from 2 µL to 10,000 µL

Depressurization value

-20 kPa ± 20 % (fixed)

Pressure tolerance inside instrument

+0.2 kPa / 10 seconds

Leak test conditions

Pressure tolerance setting: +0.1 to +20 kPa  
Monitoring time: From 0.5 seconds

Pump operation time

Up to 6 seconds

#### Options:

- Replacement tube set (Adapter × 1, attachment × 3)
- Replacement filter set (Filter unit × 1, filter element × 10)

Do you want to print the page now?

Hide some parts of page before printing



Print partially

Print the page as you see it now




Print complete page

(this box will not be printed)

Edit before printing

Hover your mouse over some part of page, a red rectangle will appear. Click it to hide the selected content of page.



Print!  Show hidden