



HELAGO-CZ, s.r.o.
Commercial Register maintained by the Regional Court in Hradec
Králové
Section C, File 17879
Kladská 1082
500 03 Hradec Králové 3
Company ID: 25 96 39 61, VAT: CZ 25963961
Phone: 495 220 229, 495 220 394
Fax: 495 220 154
GSM gate: 602 123 096
E-mail: info@helago-cz.cz
Web: http://www.helago-cz.cz

Single Channel Electronic Pipette MPA-1200

Order code: **1406.MPA1200**



Information about product price on demand

Parameters

Pipettes

Elektronické

Volume

100 - 1200 µl

Quantitative unit

ks

- Automatic operation assuring accuracy for everyone at all times
- Stress-minimized, pain-free operation using the forefinger
- Easy calibration (adjustment) by yourself - User CAL function*¹*²

- Calibration and display by either volume ($\mu\text{L}/\text{mL}$) or weight (mg)*¹
- Accurate pipetting possible with other manufacturers' tips (List of compatible tips (PDF 40KB))
- Multiple dispensing (MD) mode with automatic initial volume correction (pre-dispensing)*¹
- Mixing (MIX) mode for homogenizing solutions
- Reverse operation for accurately dispensing viscous samples
- Increased resistance to impacts from falls with padding on the head*¹
- Aspirating/dispensing speeds adjustable to 5 levels
- Storage for up to 9 preferred settings
- Quickly detachable, autoclavable lower part (121 deg C, 2 atm, 20 minutes)
- Adjustable tip ejector reach
- Long finger hook
- Power supply by either a rechargeable battery or switching (100-240V) AC adapter
- Multiple AC adapter plugs (A, BF, C, S)
- Auto power OFF function

*¹ Patent pending

*² For verification of dispensed volumes necessary for calibration, the AD-4212B-PT/ AD-4212A-PT/ FX-300iPT pipette accuracy testers (or BM series + BM-014) are useful tools