



HELAGO-CZ, s.r.o.
Commercial Register maintained by the Regional Court in Hradec Králové
Section C, File 17879
Kladská 1082
500 03 Hradec Králové 3
Company ID: 25 96 39 61, VAT: CZ 25963961
Phone: 495 220 229, 495 220 394
Fax: 495 220 154
GSM gate: 602 123 096
E-mail: info@helago-cz.cz
Web: <http://www.helago-cz.cz>

Incucell V 222 ECO
Order code: **2102.010534896**









Information about product price on demand

Parameters

Circulation	Forced
Volume	222 l
Brand, Manufacturer	BMT
Inner dimensions (w x d x h) [mm]	540 x 520 x 760
Quantitative unit	ks

- Detailed description

- Natural, precise - laboratory incubator
- 222-liter chamber volume
- Working temperature of from +10°C over ambient temperature up to 100°C
- Patented system of air flow in the chamber ensures quicker and more accurate distribution of the preset temperature in the chamber, which enables ideal heating of the samples
- High operation comfort, precise temperature regulation and short times of temperature recovery in the chamber after the door opening; the standard equipment includes inner glass door
- Model with a fan ensures quicker and more accurate distribution of temperature with small deviations; suitable for demanding and accurate tests and samples cultivation
- Standard model with a Fuzzy logic microprocessor control system; control elements situated on a foil keyboard, process information indicated on a LED display
- Wide range of optional equipment and accessories
- Output for a printer or PC connection
- Validation possibility (IQ , OQ)

Do you want to print the page now?

Hide some parts of page before printing



Print partially

Print the page as you see it now




Print complete page

(this box will not be printed)

Edit before printing

Hover your mouse over some part of page, a red rectangle will appear. Click it to hide the selected content of page.



Print!  Show hidden