



HELAGO-CZ, s.r.o.  
Commercial Register maintained by the Regional Court in Hradec  
Králové  
Section C, File 17879  
Kladská 1082  
500 03 Hradec Králové 3  
Company ID: 25 96 39 61, VAT: CZ 25963961  
Phone: 495 220 229, 495 220 394  
Fax: 495 220 154  
GSM gate: 602 123 096  
E-mail: info@helago-cz.cz  
Web: http://www.helago-cz.cz

## DT1028CH - Ultrasound bath DT 1028 CH

Order code: **2007.3266**



Information about product price on demand

Parameters

Quantitative unit

ks

### Description

- full convenience in use and display for all bath types
- easy to operate - self-explanatory symbols well recognizable parameters,
- target as well as actual values - brightly lighting LED-display
- 4-step-operation as "plug and clean"
- heater 20-80 °C, thermostatically adjustable, with LED-display for target value and actual value of temperature
- optical signal in case of excess temperature
- timer for countdown and continuous operation
- DEGAS: adjustable pulse sound level for removing unwanted air bubbles, visible at LCD-display
- automatic short-time ultrasound during heating
- up period to avoid delay in boiling - switch-off possible
- CE marked as medical device
- quality management system according to the requirements of EN ISO 13485/11.2000 in the production line
- protection class IP 33 at highest possible level through liquid-tight mains cable and housing front without openings
- SweepTec for permanent sound field oscillation ensures even and gentle cleaning
- utility patent DE 202 21 041
- automatical safety shut-down 12 hours after last keypress
- marking of filling level for safe dosage
- one-piece welded drain bow
- PZT-transducers with high efficiency

For more information about this product please download product sheet under "Files for download".