




HELAGO-CZ, s.r.o.
Commercial Register maintained by the Regional Court
in Hradec Králové
Section C, File 17879
Kladská 1082
500 03 Hradec Králové 3
Company ID: 25 96 39 61, VAT: CZ 25963961
Phone: 495 220 229, 495 220 394
Fax: 495 220 154
GSM gate: 602 123 096
E-mail: info@helago-cz.cz
Web: http://www.helago-cz.cz

NLP1400 - Breast Phantom for Mammography and Breast Tomosynthesis

Order code: **4003.NLP1400**



Information about product price on demand

Parameters

Phantoms and simulators

Quantitative unit

X-ray phantoms

ks

Breast phantom with adipose and glandular tissue. This breast phantom was designed to simulate breast imaging in mammography and breast tomosynthesis. It represents a compressed breast of 4 cm thickness that can be placed under the compression paddle. This phantom provides a realistic simulation of breast imaging. It was developed for testing and optimizing dose and image quality and for training of medical and technical staff. The phantom is manufactured from virtual data* containing adipose and glandular tissue. The models are handmade unique piece, which can differ slightly in size and design. The phantom can be provided as one-piece anthropomorphic phantom or in a sectional design. Dosimeter openings and pathologic features can be included upon request.

Size: 15x11x4cm