




HELAGO-CZ, s.r.o.  
Commercial Register maintained by the Regional  
Court in Hradec Králové  
Section C, File 17879  
Kladská 1082  
500 03 Hradec Králové 3  
Company ID: 25 96 39 61, VAT: CZ 25963961  
Phone: 495 220 229, 495 220 394  
Fax: 495 220 154  
GSM gate: 602 123 096  
E-mail: info@helago-cz.cz  
Web: http://www.helago-cz.cz

## NLP1200 - Pelvic phantom with femoral neck fracture for CT, X-ray and radiation therapy

Order code: 4003.NLP1200



Information about product price on demand

Parameters

Phantoms and simulators

Quantitative unit

X-ray phantoms

ks

This phantom is created from real patient data and is manufactured using the latest technology. Bones, vessels and soft tissue are displayed authentically with realistic CT values for all tissues at 120 kVp tube voltage in the CT. If the phantom is to be used primarily with other tube voltages (e.g. 100 kVp), the calibration of the CT values can be adjusted accordingly if required. The phantom provides realistic tissue contrasts in X-ray imaging. Air spaces are filled by a material with about -80 Hounsfield units.

The Pelvis Phantom provides a realistic simulation of a native male pelvis with a slightly dislocated femoral neck fracture on the left side. The phantom includes the entire pelvis and the fifth lumbar vertebra.

Size: 27x17x16cm