



HELAGO-CZ, s.r.o.
Commercial Register maintained by the Regional Court in
Hradec Králové
Section C, File 17879
Kladská 1082
500 03 Hradec Králové 3
Company ID: 25 96 39 61, VAT: CZ 25963961
Phone: 495 220 229, 495 220 394
Fax: 495 220 154
GSM gate: 602 123 096
E-mail: info@helago-cz.cz
Web: <http://www.helago-cz.cz>

**VET14030 - Dog pelvis phantom for
CT and x-ray**
Order code: **4003.VET14030**



Information about product price on demand

Parameters

Quantitative unit

ks

The phantom provides a realistic simulation of a dog pelvis with no contrast agent. Bones and soft tissue are displayed authentically with realistic CT values for all tissues at 120 kVp tube voltage in the CT. Air spaces are filled with a cellulose polymer composite with about -80 Hounsfield units. If the phantom is to be used primarily with other tube voltages (e.g. 100 kVp), the calibration of the CT values can be adjusted accordingly if required. The phantom also provides realistic tissue contrasts in X-ray imaging.