



HELAGO-CZ, s.r.o.  
Commercial Register maintained by the Regional Court in Hradec Králové  
Section C, File 17879  
Kladská 1082  
500 03 Hradec Králové 3  
Company ID: 25 96 39 61, VAT: CZ 25963961  
Phone: 495 220 229, 495 220 394  
Fax: 495 220 154  
GSM gate: 602 123 096  
E-mail: [info@helago-cz.cz](mailto:info@helago-cz.cz)  
Web: <http://www.helago-cz.cz>

---

**Loopster digital, rotator**  
Order code: **1305.04016000**



Cena bez DPH

1.075,00 Eur

Price with VAT

1.300,75 Eur

Parameters

Device type

Test tube shaker

Manufacturer

IKA

Maximum speed

80 rpm

Rotational Speed

0-80

Shaking motion

Rotational

Timer

YES

Maximum load	2 kg
Quantitative unit	ks

Rotator for smooth but effective mixing with adjustable speed. Ideal for biological samples, particularly blood samples. Suitable for separating and mixing processes of powdered and liquid samples in Eppendorf tubes and Greiner tubes up to 50 ml. The easy removable attachments allow simple and quick cleaning of the unit of spilled samples.

- Freely adjustable rotary angle
- Timer: Countdown, adjustable from 1 s to 99h 59min
- Counter: Display of shaking time
- Different attachments allow a variety of applications
- Usable at 4 to 50 °C in incubators
- Error code display
- Easy operation by foil keypad
- Suitable for continuous operation
- Speed range: 5 - 80 rpm