




HELAGO-CZ, s.r.o.
Commercial Register maintained by the Regional Court in Hradec Králové
Section C, File 17879
Kladská 1082
500 03 Hradec Králové 3
Company ID: 25 96 39 61, VAT: CZ 25963961
Phone: 495 220 229, 495 220 394
Fax: 495 220 154
GSM gate: 602 123 096
E-mail: info@helago-cz.cz
Web: http://www.helago-cz.cz

DW-YL90 Biomedical Freezer

Order code: **2203.DWYL90**



Information about product price on demand

Parameters

Device type	Freezer
Describtion_	2 wheels + 2 adjustable feet
Volume brutto	90 l
The temperature range in the freezer	-10 až -25°C
Number of sockets	3
Defrosting	manual
Refrigerant	R600a
Quantitative unit	ks

The high-definition digital temperature display can realize a temperature display precision of 0.1? and is convenient for observation; The high temperature/low temperature alarm, door ajar alarm, sensor failure alarm can ensure security of samples comprehensively;