

HELAGO-CZ, s.r.o.
Commercial Register maintained by the Regional Court in Hradec Králové
Section C, File 17879
Kladská 1082
500 03 Hradec Králové 3
Company ID: 25 96 39 61, VAT: CZ 25963961

Phone: 495 220 229, 495 220 394

Fax: 495 220 154 GSM gate: 602 123 096 E-mail: info@helago-cz.cz Web: http://www.helago-cz.cz P plus package for extended battery performance tests according to EUCAR hazard level 5

Order code: 2103.PKGPP



Information about product price on demand

Parameters

Quantitative unit ks

• Detailed description

Various specific battery parameters, such as state of charge, are tested in different temperature ranges. These tests are performed within a safe temperature range for the battery.

Retrofits for the instruments

- MK 115/240/720
- MKF 115/240/720
- MKT 115/240/720
- MKFT 115/240/720
- KB 115/240/400/720
- KBF 240/720
- KMF 240/720/1020

Includes

- Independent safety thermal protection class 2 switched when 120°C is reached
- Temperature range limitation adjustable on the unit controller up to 110°C
- Stainless steel pressure relief damper installed in the middle with the ceiling
- Improved door locking mechanism with reinforced hinges
- Connection for inert gas
- Temperature monitoring in the test area
- CO2 extinguishing

Do you want to print the page now? Hide some parts of page before printing



Print partially

Print the page as you see it now



Print complete page

(this box will not be printed)

Edit before printing

Hover your mouse over some part of page, a red rectangle will appear. Click it to hide the selected content of page.

